

BS EN 45544-1:2015



BSI Standards Publication

Workplace atmospheres — Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours

Part 1: General requirements and test
methods

bsi.

...making excellence a habit.™

National foreword

This British Standard is the UK implementation of EN 45544-1:2015. It supersedes BS EN 45544-1:2000 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee EXL/31/1, Gas detectors.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015. Published by BSI Standards Limited 2015

ISBN 978 0 580 81573 7

ICS 13.040.30; 13.320

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 28 February 2015.

Amendments issued since publication

Date	Text affected
------	---------------

English Version

Workplace atmospheres - Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours - Part 1: General requirements and test methods

Atmosphères des lieux de travail - Appareillage électrique utilisé pour la détection directe des vapeurs et gaz toxiques et le mesurage direct de leur concentration - Partie 1: Exigences générales et méthodes d'essai


Arbeitsplatzatmosphäre - Elektrische Geräte für die direkte Detektion und direkte Konzentrationsmessung toxischer Gase und Dämpfe - Teil 1: Allgemeine Anforderungen und Prüfverfahren

This European Standard was approved by CENELEC on 24 November 2014. CEN and CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN and CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN and CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN and CENELEC members are the national standards bodies and national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword	4
Introduction.....	5
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions	7
4 General requirements	14
4.1 Introduction.....	14
4.2 Construction	14
4.2.1 General	14
4.2.2 Indicating devices	14
4.2.3 Alarm output functions	16
4.3 Fault signals.....	16
4.4 Adjustments.....	16
4.5 Battery-powered apparatus.....	17
4.6 Gas detection transmitter for use with separate control units.....	17
4.7 Separate control units for use with gas detection transmitters.....	17
4.8 Apparatus using software and/or digital technologies	17
4.9 Labelling and marking	17
4.9.1 General	17
4.9.2 Identification of the gas to be detected.....	18
4.10 Instruction manual	18
5 Test methods	21
5.1 General requirements for tests	21
5.2 Samples and sequence of tests.....	21
5.2.1 Test samples	21
5.2.2 Sequence.....	21
5.2.3 Gas detection transmitters.....	21
5.2.4 Separate control units.....	21
5.2.5 Test of compliance with general requirements	22
5.2.6 Apparatus with selectable range	22
5.2.7 Preparation of apparatus before each particular test.....	22
5.2.8 Mask for calibration and test.....	23
5.3 Normal conditions for test.....	23
5.3.1 General	23
5.3.2 Test gas(es).....	24
5.3.3 Flow rate for test gases	24
5.3.4 Power supply	24
5.3.5 Temperature.....	24
5.3.6 Pressure	24
5.3.7 Humidity	24
5.3.8 Stabilization	25
5.3.9 Orientation	25
5.3.10 Communications options	25
5.3.11 Gas detection apparatus as part of systems.....	25

5.4	Tests	25
5.4.1	General	25
5.4.2	Unpowered storage	26
5.4.3	Measurement of deviations	26
5.4.4	Mechanical tests	27
5.4.5	Environmental tests	28
5.4.6	Performance tests	30
5.4.7	Electrical tests	34
5.4.8	Stability	35
5.4.9	Verification of software and digital components	35
6	Uncertainty of measurement and lower limit of measurement	35
6.1	General	35
6.2	Method of calculation of uncertainty of measurement	36
6.2.1	General	36
6.2.2	Sources of uncertainty	37
6.2.3	Calculation of relative expanded uncertainty	42
6.3	Method of calculation of lower limit of measurement	42
7	Test report	44
Annex A (normative) Standard test gas volume fractions for EN 45544-2 apparatus		45
Annex B (informative) Sequence of tests		47
Annex C (informative) Example of a test chamber		48
Annex D (informative) Table of significant changes in comparison to EN 45544-1:1999		49

Foreword

This document (EN 45544-1:2015) has been prepared by CEN/CENELEC Joint Working Group Continuous Measuring Instruments (JWG CMI).

The following dates are fixed:

- latest date by which this document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2015-11-24
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2017-11-24

This document supersedes EN 45544-1:1999.

Introduction

National laws and regulations based on European Directives require the assessment of the potential exposure of a worker to chemical agents in workplace atmospheres.

EN 45544, *Workplace atmospheres – Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours*, consists of the following parts:

- *Part 1: General requirements and test methods;*
- *Part 2: Performance requirements for apparatus used for exposure measurement;*
- *Part 3: Performance requirements for apparatus used for general gas detection;*
- *Part 4: Guide for selection, installation, use and maintenance.*

EN 45544 series is based on EN 482 which specifies general performance requirements for procedures for measuring the concentration of chemical agents in workplace atmospheres. These performance requirements are intended to apply under environmental conditions present at the workplace. However, because a wide range of environmental conditions are encountered in practice, this document specifies requirements that have to be fulfilled by measuring procedures when tested under prescribed laboratory conditions.

EN 45544-2 details the performance requirements outlined in EN 482 specifically for electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours intended for exposure measurement.

EN 45544-3 details the performance requirements for general gas detection apparatus, e.g. safety warning, leak detection. The measuring range will be defined by the manufacturer. In general, the requirements for accuracy will be higher for EN 45544-2 apparatus than for EN 45544-3 apparatus.

The same apparatus may be used for applications covered by EN 45544-2 and EN 45544-3.

EN 45544 series will help manufacturers, test laboratories and users of apparatus to adopt a consistent approach to, and provide a framework for, the assessment of performance criteria. It is the manufacturer's primary responsibility to ensure that the apparatus meets the requirements laid down in this European Standard, including environmental influences, which can be expected to affect performance.

For a given measurement task, the range over which the requirements for the relative expanded uncertainty have to be met depends on the limit value. However, for most chemical agents the limit values have not been harmonized at the European level. Therefore, it was decided to use a reference value (standard test gas concentration) instead of the limit value for the performance tests. The list of standard test gas volume fractions is given in Annex A. The values chosen are equal to or close to the limit values used in different European countries but are intended to be used only for type testing apparatus without any legal implications.

Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours generate readings in clean air (zero readings), which vary with environmental conditions and time. This standard therefore includes test methods and requirements for acceptable variations in zero readings which are additional to the general requirements of EN 482.

1 Scope

This European Standard specifies general requirements and test methods for the determination of the performance characteristics of personal, portable, transportable and fixed, continuous duty electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours in workplace atmospheres.

This European Standard is applicable to apparatus whose primary purpose is to provide an indication, alarm and/or other output function to give a warning of the presence of a toxic gas or vapour in the atmosphere and, in some cases, to initiate automatic or manual protective actions. It is applicable to apparatus in which the sensor automatically generates an electrical signal when gas is present.

This European Standard is not applicable to apparatus:

- used for the measurement of oxygen;
- used only in laboratories for analysis or measurement;
- used only for process measurement purposes;
- used in car parks or tunnels;
- used in the domestic environment;
- used in environmental air pollution monitoring;
- used for the measurement of combustible gases and vapours related to the risk of explosion.

It also does not apply to open-path (line of sight) area monitors.

For apparatus used for sensing the presence of multiple gases this standard applies only to the detection of toxic gas or vapour.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 482, *Workplace exposure - General requirements for the performance of procedures for the measurement of chemical agents*

EN 45544-2:2015, *Workplace atmospheres - Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours - Part 2: Performance requirements for apparatus used for exposure measurement*

EN 45544-3:2015, *Workplace atmospheres - Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours - Part 3: Performance requirements for apparatus used for general gas detection*

EN 50270, *Electromagnetic compatibility - Electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen*