



BSI Standards Publication

Needle-based injection systems for medical use — Requirements and test methods

Part 4: Needle-based injection systems containing electronics

National foreword

This British Standard is the UK implementation of EN ISO 11608-4:2022. It is identical to [ISO 11608-4:2022](#). It supersedes [BS EN ISO 11608-4:2007](#), which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/84, Catheters and syringes.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2022
Published by BSI Standards Limited 2022

ISBN 978 0 539 03130 0

ICS 11.040.25

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 July 2022.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

EUROPEAN STANDARD

EN ISO 11608-4

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2022

ICS 11.040.25

Supersedes EN ISO 11608-4:2007

English Version

Needle-based injection systems for medical use
- Requirements and test methods - Part 4: Needle-
based injection systems containing electronics (ISO
11608-4:2022)

Systèmes d'injection à aiguille pour usage
médical - Exigences et méthodes d'essai - Partie
4: Systèmes d'injection à aiguille contenant
de l'électronique (ISO 11608-4:2022)

Kanülenbasierte Injektionssysteme zur
medizinischen Verwendung - Anforderungen
und Prüfverfahren - Teil 4: Kanülenbasierte
Injektionssysteme, die elektronische
Bauteile enthalten (ISO 11608-4:2022)

This European Standard was approved by CEN on 2 January 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 11608-4:2022) has been prepared by Technical Committee ISO/TC 84 "Devices for administration of medicinal products and catheters" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2022, and conflicting national standards shall be withdrawn at the latest by November 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11608-4:2007.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of [ISO 11608-4:2022](#) has been approved by CEN as EN ISO 11608-4:2022 without any modification.

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	3
4 Abbreviated terms	11
5 General requirements	11
5.1 Conditions for application of NIS-E.....	11
5.2 General design requirements.....	11
5.3 Risk approach and usability engineering.....	12
6 General requirements for testing	13
6.1 Type tests.....	13
6.2 Number of samples.....	13
6.3 Ambient temperature, humidity, atmospheric pressure.....	18
7 Identification and marking of NIS-E	18
8 Protection against electrical hazards	18
8.1 General.....	18
8.2 Humidity preconditioning treatment.....	18
8.3 Requirements and test methods.....	19
8.3.1 General.....	19
8.3.2 Applied parts.....	19
8.3.3 Requirements related to power sources.....	21
8.3.4 Limitation of current for accessible parts and applied parts.....	22
8.4 Separation of parts (Type X and Type Y).....	22
8.4.1 Means of protection (MOP).....	22
8.4.2 Separation of patient connection.....	23
8.4.3 Maximum mains voltage.....	24
8.4.4 Working voltage.....	24
8.5 Patient leakage current and touch current (Type X and Type Y NIS-E).....	25
8.5.1 General.....	25
8.5.2 Measurement of patient leakage current.....	29
8.5.3 Measurement of touch current.....	33
8.6 Insulation (Type X and Type Y).....	33
8.6.1 General.....	33
8.6.2 Distance through solid insulation or use of thin sheet material.....	34
8.6.3 Dielectric strength.....	34
8.7 Insulation other than wire insulation.....	35
8.7.1 Mechanical strength and resistance to heat.....	35
8.8 Creepage distances and air clearances (Type X and Type Y NIS-E).....	36
8.8.1 General.....	36
8.9 Specific hazardous situations.....	36
8.9.1 General.....	36
8.9.2 Emissions, deformation of enclosure or exceeding maximum temperature.....	36
8.9.3 Exceeding leakage current or voltage limits.....	38
8.9.4 Specific mechanical hazards.....	39
8.10 Single fault conditions (Type X and Type Y).....	39
8.10.1 General.....	39
8.10.2 Failure of thermostats and temperature limiting devices.....	39
8.10.3 Leakage of liquid from batteries.....	39
8.10.4 Locking of moving parts.....	39
8.10.5 Additional test criteria for motor-operated NIS-E.....	39

8.10.6	NIS-E intended for used in conjunction with oxygen rich environments.....	40
8.10.7	Power supply (Type Y)	40
8.11	Pre-conditioning for the influence of fluid leakage.....	41
9	Electromagnetic compatibility (EMC)	41
9.1	General requirements.....	41
9.1.1	Risk approach process for NIS-E.....	41
9.1.2	Non-medical electrical equipment used with NIS-E	42
9.1.3	General test conditions.....	42
9.2	NIS-E identification, marking and documents.....	47
9.2.1	Instruction for use in relation to EMC.....	47
9.2.2	Documentation of the tests.....	47
9.3	Electromagnetic emissions requirements for NIS-E	48
9.3.1	Protection of radio services and other equipment.....	48
9.3.2	Protection of the public mains network.....	48
9.3.3	Emissions requirements summary (Type X and Type Y)	49
9.4	Electromagnetic immunity requirements for NIS-E.....	49
9.4.1	General.....	49
9.4.2	Operating modes.....	51
9.4.3	Non-medical electrical equipment.....	51
9.4.4	Immunity test levels.....	51
9.4.5	Immunity to proximity fields from RF wireless communications equipment	56
9.4.6	Immunity to proximity magnetic fields in the frequency range 9 kHz to 13,56 MHz.....	58
10	Protection against mechanical hazards.....	58
10.1	General.....	58
10.2	Shock.....	58
10.3	Vibration.....	58
10.3.1	Sinusoidal vibration.....	58
10.3.2	Random vibration.....	58
10.4	Impact of OBDS enclosures	59
10.5	Push.....	59
11	Programmable NIS-E.....	59
Annex A (informative) Identification of immunity pass/fail criteria.....		60
Annex B (informative) Rationale for using 240 V for testing some requirements		62
Bibliography		63

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 84, *Devices for administration of medicinal products and catheters*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition ([ISO 11608-4:2006](http://www.iso.org/iso/11608-4:2006)), which has been technically revised.

The main changes are as follows:

- this document has been revised in its entirety to include requirements from the IEC 60601 series that pertain to hand-held medical injectors.

A list of all parts in the ISO 11608 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Needle-based injection systems, including on-body delivery systems (OBDSs), containing electronics with or without software, are primarily intended to administer medicinal products to humans. Performance requirements regarding essential electrotechnical aspects have been selected with the intention not to restrict the Electronic Needle-based Injection System (NIS-E) design unnecessarily when applying the document.

The first edition of this document was limited to pen-injectors with electromechanical drive. Pen-injectors only equipped with electronics were covered in [ISO 11608-1](#).

Materials used for construction are not specified in this document, as their selection will depend on the design, the intended use and the process of manufacture used by individual manufacturers.

There are other international and national standards and guidance publications and, in some countries, national regulations that are applicable to medical devices and pharmaceuticals. This document is applicable to NIS-E and specifies relevant aspects of IEC 60601-1:2005+AMD1:2012+AMD2:2020, IEC 60601-1-2:2014+AMD1:2020 and IEC 60601-1-11:2015+AMD1:2020 for this particular device type.

This document does not specify non-electrotechnical requirements and test methods for NISs when specified by [ISO 11608-1](#).

Developers and manufacturers of NIS-Es are encouraged to investigate and determine whether there are any other requirements relevant to the safety or marketability of their NIS-Es. For example, this document should be used in conjunction with [IEC 60601-1](#), [IEC 60601-1-2](#) and [IEC 60601-1-11](#). A risk-based approach is expected to be applied during the design, development, and manufacture of the product. Given the specific medicinal product intended use and environment, this might result in product-specific requirements and test methods that differ from what is outlined in this document.

This document is intended to be used for type testing (testing of the development result) of NIS-E. It is not intended to be used for batch release testing.

This document introduces the notion of Type X NIS-E and Type Y NIS-E. Type X NIS-E is a device type without any physical cabled connection to other devices. Type Y NIS-E has such connections. The electrical requirements in this document for Type X NIS-E is a subset of the requirements for Type Y NIS-E.

Needle-based injection systems for medical use — Requirements and test methods —

Part 4: Needle-based injection systems containing electronics

1 Scope

This document specifies requirements and test methods for needle-based injection systems (NISs) containing electronics with or without software (NIS-Es).

The needle-based injection system containing electronics can be single use or reusable and can be operated with or without electrical/conductive connections to other devices. The system is intended to deliver medication to a patient by self-administration or by administration by one other operator (e.g. caregiver or health care provider).

This document applies to electronic accessories that are intended to be physically connected to a NIS or NIS-E according to the NIS/NIS-E intended use.

This document also applies to electronic accessories that are intended to have electrical/conductive connections to a NIS or NIS-E according to the NIS/NIS-E intended use.

This document does not specify requirements for software in programmable NIS-E.

NOTE IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 14 addresses software life cycle processes.

This document does not specify requirements for cybersecurity.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

[CISPR 11](#), *Industrial, scientific and medical equipment – Radio-frequency disturbance characteristics – Limits and methods of measurement*

[CISPR 32](#), *Electromagnetic compatibility of multimedia equipment — Emission requirements*

[ISO 11608-1:2022](#), *Needle-based injection systems for medical use — Requirements and test methods — Part 1: Needle-based injection systems*

[ISO 7137](#), *Aircraft — Environmental conditions and test procedures for airborne equipment*

[ISO 14971:2019](#), *Medical devices — Application of risk management to medical devices*

[IEC 60086-4](#), *Primary batteries — Part 4: Safety of lithium batteries*

[IEC 60068-2-64](#), *Environmental testing — Part 2-64: Tests — Test Fh: Vibration, broadband random and guidance*

[IEC 60529](#), *Degrees of protection provided by enclosures (IP Code)*

IEC 60601-1:2005+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*