

---

---

**Biological evaluation of medical  
devices —**

Part 10:  
**Tests for irritation and skin sensitization**

*Évaluation biologique des dispositifs médicaux —*

*Partie 10: Essais d'irritation et de sensibilisation cutanée*



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

Foreword .....	iv
Introduction.....	vi
<b>1 Scope .....</b>	<b>1</b>
<b>2 Normative references .....</b>	<b>1</b>
<b>3 Terms and definitions .....</b>	<b>2</b>
<b>4 General principles — Step-wise approach .....</b>	<b>4</b>
<b>5 Pretest considerations .....</b>	<b>4</b>
<b>5.1 General .....</b>	<b>4</b>
<b>5.2 Types of material .....</b>	<b>5</b>
<b>5.3 Information on chemical composition .....</b>	<b>5</b>
<b>6 Irritation tests .....</b>	<b>6</b>
<b>6.1 <i>In vitro</i> irritation tests .....</b>	<b>6</b>
<b>6.2 <i>In vivo</i> irritation tests — Factors to be considered in design and selection of <i>in vivo</i> tests .....</b>	<b>6</b>
<b>6.3 Animal irritation test .....</b>	<b>7</b>
<b>6.4 Animal intracutaneous (intra-dermal) reactivity test .....</b>	<b>11</b>
<b>6.5 Human skin irritation test .....</b>	<b>14</b>
<b>7 Skin sensitization tests .....</b>	<b>15</b>
<b>7.1 Choice of test methods .....</b>	<b>15</b>
<b>7.2 Murine Local Lymph Node Assay (LLNA) .....</b>	<b>15</b>
<b>7.3 Guinea pig assays for the detection of skin sensitization .....</b>	<b>18</b>
<b>7.4 Important factors affecting the outcome of the test .....</b>	<b>19</b>
<b>7.5 Guinea pig maximization test (GPMT) .....</b>	<b>20</b>
<b>7.6 Closed-patch test (Buehler test) .....</b>	<b>23</b>
<b>8 Key factors in interpretation of test results .....</b>	<b>26</b>
<b>Annex A (normative) Preparation of materials for irritation/sensitization testing .....</b>	<b>27</b>
<b>Annex B (normative) Special irritation tests .....</b>	<b>29</b>
<b>Annex C (normative) Human skin irritation test .....</b>	<b>44</b>
<b>Annex D (informative) <i>In vitro</i> tests for skin irritation .....</b>	<b>48</b>
<b>Annex E (informative) Method for the preparation of extracts from polymeric test materials .....</b>	<b>54</b>
<b>Annex F (informative) Background information .....</b>	<b>57</b>
<b>Bibliography .....</b>	<b>61</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 10993-10 was prepared by Technical Committee ISO/TC 194, *Biological evaluation of medical devices*.

This third edition cancels and replaces the second edition (ISO 10993-10:2002), which has been technically revised.

ISO 10993 consists of the following parts, under the general title *Biological evaluation of medical devices*:

- *Part 1: Evaluation and testing within a risk management process*
- *Part 2: Animal welfare requirements*
- *Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity*
- *Part 4: Selection of tests for interactions with blood*
- *Part 5: Tests for in vitro cytotoxicity*
- *Part 6: Tests for local effects after implantation*
- *Part 7: Ethylene oxide sterilization residuals*
- *Part 9: Framework for identification and quantification of potential degradation products*
- *Part 10: Tests for irritation and skin sensitization*
- *Part 11: Tests for systemic toxicity*
- *Part 12: Sample preparation and reference materials*
- *Part 13: Identification and quantification of degradation products from polymeric medical devices*
- *Part 14: Identification and quantification of degradation products from ceramics*
- *Part 15: Identification and quantification of degradation products from metals and alloys*

- *Part 16: Toxicokinetic study design for degradation products and leachables*
- *Part 17: Establishment of allowable limits for leachable substances*
- *Part 18: Chemical characterization of materials*
- *Part 19: Physico-chemical, morphological and topographical characterization of materials* [Technical Specification]
- *Part 20: Principles and methods for immunotoxicology testing of medical devices* [Technical Specification]

## Introduction

This part of ISO 10993 assesses possible contact hazards from chemicals released from medical devices, which may produce skin and mucosal irritation, eye irritation or skin sensitization.

Some materials that are included in medical devices have been tested, and their skin or mucosal irritation or sensitization potential has been documented. Other materials and their chemical components have not been tested and may induce adverse effects when in contact with human tissue. The manufacturer is thus obliged to evaluate each device for potential adverse effects prior to marketing.

Traditionally, small animal tests are performed prior to testing on humans to help predict human response. More recently, *in vitro* tests as well as human tests have been added as adjuncts or alternatives. Despite progress and considerable effort in this direction, a review of findings suggests that currently no satisfactory *in vitro* test has been devised to eliminate the requirement for *in vivo* testing. Where appropriate, the preliminary use of *in vitro* methods is encouraged for screening purposes prior to animal testing. In order to reduce the number of animals used, this part of ISO 10993 presents a step-wise approach, with review and analysis of test results at each stage. An animal test is usually required prior to human testing.

It is intended that these studies be conducted using Good Laboratory Practice and comply with regulations related to animal welfare. Statistical analysis of data is recommended and should be used whenever appropriate.

This part of ISO 10993 is intended for use by professionals, appropriately qualified by training and experience, who are able to interpret its requirements and judge the outcomes of the evaluation for each medical device, taking into consideration all the factors relevant to the device, its intended use and the current knowledge of the medical device provided by review of the scientific literature and previous clinical experience.

The tests included in this part of ISO 10993 are important tools for the development of safe products, provided that these are executed and interpreted by trained personnel.

This part of ISO 10993 is based on numerous standards and guidelines, including OECD Guidelines, U.S. Pharmacopoeia and the European Pharmacopoeia. It is intended to be the basic document for the selection and conduct of tests enabling evaluation of irritation and dermal sensitization responses relevant to safety of medical materials and devices.

# Biological evaluation of medical devices —

## Part 10:

## Tests for irritation and skin sensitization

### 1 Scope

This part of ISO 10993 describes the procedure for the assessment of medical devices and their constituent materials with regard to their potential to produce irritation and skin sensitization.

This part of ISO 10993 includes:

- a) pretest considerations for irritation, including *in silico* and *in vitro* methods for dermal exposure;
- b) details of *in vivo* (irritation and sensitization) test procedures;
- c) key factors for the interpretation of the results.

Instructions are given in Annex A for the preparation of materials specifically in relation to the above tests. In Annex B several special irritation tests are described for application of medical devices in areas other than skin.

### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1:2009, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 10993-2, *Biological evaluation of medical devices — Part 2: Animal welfare requirements*

ISO 10993-9, *Biological evaluation of medical devices — Part 9: Framework for identification and quantification of potential degradation products*

ISO 10993-12, *Biological evaluation of medical devices — Part 12: Sample preparation and reference materials*

ISO 10993-13, *Biological evaluation of medical devices — Part 13: Identification and quantification of degradation products from polymeric medical devices*

ISO 10993-14, *Biological evaluation of medical devices — Part 14: Identification and quantification of degradation products from ceramics*

ISO 10993-15, *Biological evaluation of medical devices — Part 15: Identification and quantification of degradation products from metals and alloys*

ISO 10993-18, *Biological evaluation of medical devices — Part 18: Chemical characterization of materials*

ISO 14155-1, *Clinical investigation of medical devices for human subjects — Part 1: General requirements*

ISO 14155-2, *Clinical investigation of medical devices for human subjects — Part 2: Clinical investigation plans*