

INTERNATIONAL
STANDARD

ISO
18321

First edition
2015-05-15

Essential oils — Determination of peroxide value

Huiles essentielles — Détermination de l'indice de peroxyde



Reference number
ISO 18321:2015(E)

© ISO 2015



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	1
6 Apparatus	2
7 Sampling	2
8 Preparation of test sample	2
9 Procedure	2
9.1 Test portion.....	2
9.2 Determination.....	3
9.3 Automated potentiometric titration.....	3
10 Expression of results	3
10.1 Calculation.....	3
10.2 Conversion from mmol/l to meq/kg.....	3
11 Precision	4
11.1 Repeatability.....	4
11.2 Reproducibility.....	4
Bibliography	5

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: [Foreword — Supplementary information](#).

The committee responsible for this document is ISO/TC 54, *Essential oils*.

Essential oils — Determination of peroxide value

1 Scope

This International Standard specifies a method for the determination of the peroxide value in an essential oil. The peroxide value is a measure of the oxidation present.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 212, *Essential oils — Sampling*

ISO 356, *Essential oils — Preparation of test samples*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

For the purpose of this document, the following terms and definitions apply.

3.1

peroxide value

Ip

number that expresses, in millimoles (or milliequivalents), the quantity of peroxide contained in 1 000 ml of the substance

4 Principle

It is a redox titration of the iodometry type. The iodide ions added to the essential oil oxidize when reacting with peroxides, obtaining iodine which is titrated with thiosulphate. It can be carried out by volumetric or potentiometric titration.

Potentiometric titration is particularly recommended for highly coloured essential oils for which are difficult to appreciate the end point of the coloured indicator (e.g. vetiver essential oil).

5 Reagents

During the analysis, only reagents of recognized analytical grade and reverse osmosis or distilled or deionized water of Grade 3, as defined in ISO 3696, should be used.

5.1 Trichloromethane (chloroform), 99 % (volume/fraction), or cyclohexane, 99,5 % (volume/fraction) for laboratories with restrictions on the use of chloroform.

5.2 Glacial acetic acid, 99,5 % (volume/fraction). Degassed in an ultrasonic bath or by purging with a current pure and dry inert gas (carbon dioxide or nitrogen).

5.3 Potassium iodide, saturated solution of potassium iodide in deionised water, freshly prepared. The solution must remain saturated (undissolved crystals must be present). The solution has to be kept protected from light.