



BSI Standards Publication

## **Sterilization of health care products - Biological and chemical indicators - Test equipment**

---

## National foreword

This British Standard is the UK implementation of EN ISO 18472:2018. It is identical to ISO 18472:2018. It supersedes BS EN ISO 18472:2006, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/198, Sterilization and Associated Equipment and Processes.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2018  
Published by BSI Standards Limited 2018

ISBN 978 0 580 91730 1

ICS 11.080.01

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2018.

### Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

---

EUROPEAN STANDARD

**EN ISO 18472**

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2018

ICS 11.080.01

Supersedes EN ISO 18472:2006

English Version

## Sterilization of health care products - Biological and chemical indicators - Test equipment (ISO 18472:2018)

Stérilisation des produits de santé - Indicateurs biologiques et chimiques - Appareillage d'essai (ISO 18472:2018)

Sterilisation von Produkten für die Gesundheitsfürsorge - Biologische und chemische Indikatoren - Prüfausrüstung (ISO 18472:2018)

This European Standard was approved by CEN on 9 May 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

## European foreword

This document (EN ISO 18472:2018) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 102 "Sterilizers and associated equipment for processing of medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2019, and conflicting national standards shall be withdrawn at the latest by March 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 18472:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Endorsement notice

The text of ISO 18472:2018 has been approved by CEN as EN ISO 18472:2018 without any modification.

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>2</b>
<b>4 Performance requirements for resistometers</b> .....	<b>4</b>
4.1 Intended use.....	4
4.2 Test methods.....	4
4.3 Air leakage test.....	4
4.4 Steam resistometer performance requirements .....	5
4.4.1 Measurement accuracy .....	5
4.4.2 Data .....	5
4.4.3 Process control.....	5
4.4.4 General steam resistometer requirements .....	6
4.4.5 Air leakage test.....	7
4.4.6 Operation of steam resistometer.....	7
4.5 Ethylene oxide gas resistometer performance requirements.....	7
4.5.1 Measurement accuracy .....	7
4.5.2 Data .....	8
4.5.3 Process control.....	8
4.5.4 General ethylene oxide gas resistometer requirements .....	9
4.5.5 Air leakage test.....	10
4.5.6 Operation of ethylene oxide gas resistometer.....	10
4.6 Dry heat (heated air) resistometer performance requirements.....	10
4.6.1 Measurement accuracy .....	10
4.6.2 Data .....	11
4.6.3 Process control.....	11
4.6.4 General dry heat (heated air) resistometer requirements.....	12
4.6.5 Operation of dry heat (heated air) resistometer.....	12
4.7 Vaporized hydrogen peroxide resistometer performance requirements .....	13
4.7.1 Measurement accuracy .....	13
4.7.2 Recording interval.....	13
4.7.3 Process control.....	13
4.7.4 General vaporized hydrogen peroxide resistometer requirements.....	14
4.7.5 Air leakage test.....	15
4.7.6 Operation of vaporized hydrogen peroxide resistometer.....	15
<b>5 Calibration</b> .....	<b>15</b>
<b>Annex A (informative) Additional performance characterization — Steam</b> .....	<b>16</b>
<b>Annex B (informative) Additional performance characterization — Ethylene oxide gas</b> .....	<b>19</b>
<b>Annex C (informative) Additional performance characterization — Dry heat</b> .....	<b>22</b>
<b>Annex D (informative) Resistometer documentation and derivations</b> .....	<b>24</b>
<b>Bibliography</b> .....	<b>30</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This second edition cancels and replaces the first edition (ISO 18472:2006), which has been technically revised.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

To test the performance of biological and chemical indicators, specific test equipment is required. This document specifies the performance requirements for the test equipment to be used to establish the response of biological and chemical indicators to critical process variables. This document does not apply to test equipment for indicators used in irradiation, isolator/room biodecontamination (at atmospheric pressure), or low temperature steam and formaldehyde processes.

Resistometers constitute test equipment designed to create precise and repeatable sterilizing environments, allowing the evaluation of their effect on biological inactivation kinetics, chemical reactions, material degradation and product bioburden. Resistometers allow precise variation of the environmental conditions and cycle sequences in order to produce controlled physical studies. When used with the defined test methods given in the appropriate parts of ISO 11138 for biological indicators and ISO 11140 for chemical indicators, the results of these studies can be used to demonstrate conformance of biological indicators and chemical indicators to these standards.

Resistometers differ from conventional sterilizers. Instrumentation selection and control requirements for resistometers are based upon mathematical models in which rates of reaction, measurement accuracy and process control requirements are evaluated to quantify the effects induced by test equipment-controlled variables. The requirements for accurate measurement, precise control, and rapid rates of change approach limits of commercially available process control and calibration instrumentation measurement accuracy. The measurement and control requirements often prohibit practical validation of a resistometer using procedures that might be employed in a conventional heat or chemical sterilization system. Resistometers are considered test equipment rather than sterilizers; therefore, an understanding of instrumentation and process design is critical in clarifying requirements on precision and measurement accuracy. Practical design takes the following into consideration:

- achievable measurement and control;
- acceptable equipment induced variation in test results;
- economic design (utilizing tight process controls only where required);
- test method correlation with intended use;
- historical knowledge applied to test procedures and an understanding of micro-environmental physical phenomena;
- testing and analysis alternatives, when accurate quantitative determinations exceed physical measurement/control limits.

# Sterilization of health care products — Biological and chemical indicators — Test equipment

## 1 Scope

This document specifies the requirements for test equipment to be used to:

- test biological indicators for steam, ethylene oxide gas and dry heat sterilization processes for conformity to the requirements given in ISO 11138 series;
- test chemical indicators for steam, ethylene oxide gas, dry heat and vaporized hydrogen peroxide sterilization processes for conformity to the requirements given in ISO 11140-1:2014.

This document also provides informative methods useful in characterizing the performance of biological and chemical indicators for intended use and for routine quality control testing.

This document does not specify requirements for test equipment for processes specifically for testing chemical and biological indicators intended to monitor isolator and room biodecontamination processes at atmospheric pressure.

ISO 11138-2:2017, ISO 11138-3:2017, ISO 11138-4:2017 and ISO 11140-1:2014 require the use of resistometers specified in this document, and these resistometers are used in conjunction with the test methods specified in the appropriate parts of ISO 11138 series and ISO 11140 series.

Resistometers for low temperature steam and formaldehyde indicators are not included in this document. Test methods using laboratory apparatus for low temperature steam and formaldehyde are included in ISO 11138-5:2017.

Test equipment for testing Type 2 (e.g. Bowie Dick) chemical indicators are specified in ISO 11140-3:2007, ISO 11140-4:2007, and ISO 11140-5:2007.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11138-1:2017, *Sterilization of health care products — Biological indicators — Part 1: General requirements*

ISO 11138-2:2017, *Sterilization of health care products — Biological indicators — Part 2: Biological indicators for ethylene oxide sterilization processes*

ISO 11138-3:2017, *Sterilization of health care products — Biological indicators — Part 3: Biological indicators for moist heat sterilization processes*

ISO 11138-4:2017, *Sterilization of health care products — Biological indicators — Part 4: Biological indicators for dry heat sterilization processes*

ISO 11138-5:2017, *Sterilization of health care products — Biological indicators — Part 5: Biological indicators for low-temperature steam and formaldehyde sterilization processes*

ISO 11140-1:2014, *Sterilization of health care products — Chemical indicators — Part 1: General requirements*