



BSI Standards Publication

Specification for radio disturbance and immunity measuring apparatus and methods

Part 1-4: Radio disturbance and immunity measuring apparatus —
Antennas and test sites for radiated disturbance measurements

National foreword

This British Standard is the UK implementation of EN IEC 55016-1-4:2019+A2:2023. It is identical to CISPR 16-1-4:2019, incorporating amendment 1:2020 and amendment 2:2023. It supersedes BS EN IEC 55016-1-4:2019, which will be withdrawn on 20 July 2023.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to CISPR text carry the number of the CISPR amendment. For example, text altered by CISPR amendment 1 is indicated by **A1** **A1**.

The text of CISPR amendment 2:2023 has been provided in its entirety at the end of this document. BSI's policy of providing consolidated content remains unchanged; however, in the interest of expediency, in this instance BSI have chosen to collate the relevant content at the end of this document.

The UK participation in its preparation was entrusted to Technical Committee GEL/210, EMC - Policy committee.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 03588 9

ICS 17.220.20; 33.100.10; 33.100.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 March 2019..

Amendments/corrigenda issued since publication

Date	Text affected
31 August 2020	Implementation of CISPR amendment 1:2020 with CENELEC endorsement A1:2020

Date	Text affected
30 June 2023	Implementation of CISPR amendment 2:2023 with CENELEC endorsement A2:2023

EUROPEAN STANDARD

EN IEC 55016-1-4:2019ŽA2

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2023

ICS 33.100.10; 33.100.20

English Version

**Specification for radio disturbance and immunity measuring apparatus and methods - Part 1-4: Radio disturbance and immunity measuring apparatus - Antennas and test sites for radiated disturbance measurements
(CISPR 16-1-4:2019)**

Spécifications des méthodes et des appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques - Partie 1-4: Appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques - Antennes et emplacements d'essai pour les mesures des perturbations rayonnées
(CISPR 16-1-4:2019)

Anforderungen an Geräte und Einrichtungen sowie Festlegung der Verfahren zur Messung der hochfrequenten Störaussendung (Funkstörungen) und Störfestigkeit - Teil 1-4: Geräte und Einrichtungen zur Messung der hochfrequenten Störaussendung (Funkstörungen) und Störfestigkeit - Antennen und Messplätze für Messungen der gestrahlten Störaussendung
(CISPR 16-1-4:2019)

This European Standard was approved by CENELEC on 2019-02-12. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document CIS/A/1262/FDIS, future edition 4 of CISPR 16-1-4, prepared by CISPR SC A "Radio-interference measurements and statistical methods" of CISPR "International special committee on radio interference" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 55016-1-4:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2019-11-12
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-02-12

This document supersedes EN 55016-1-4:2010.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard CISPR 16-1-4:2019 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 61169-8	NOTE	Harmonized as EN 61169-8
IEC 61000-4-20	NOTE	Harmonized as EN 61000-4-20

Foreword to amendment A1

The text of document CIS/A/1316/FDIS, future CISPR 16-1-4/A1, prepared by CISPR SC A "Radio-interference measurements and statistical methods" of CISPR "International special committee on radio interference" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 55016-1-4:2019/A1:2020.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2021-04-20
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2023-07-20

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard CISPR 16-1-4:2019/A1:2020 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

CISPR 11	NOTE	Harmonized as EN 55011
----------	------	------------------------

European foreword

The text of document CIS/A/1389/FDIS, future CISPR 16-1-4/AMD2, prepared by CISPR SC A "Radio-interference measurements and statistical methods" of CISPR "International special committee on radio interference" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 55016-1-4:2019/A2:2023.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2024-02-18
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2026-05-18

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a Standardization Request given to CENELEC by the European Commission and the European Free Trade Association.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

Endorsement notice

The text of the International Standard CISPR 16-1-4:2019/AMD2:2023 was approved by CENELEC as a European Standard without any modification.

EN IEC 55016-1-4:2019+A2:2023 (E)

Annex ZA




(normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.



NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

	<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
	CISPR 16-1-1	2019	Specification for radio disturbance and immunity measuring apparatus and methods - Part 1-1: Radio disturbance and immunity measuring apparatus - Measuring apparatus	EN IEC 55016-1-1	2019
	CISPR 16-1-5	2014	Specification for radio disturbance and immunity measuring apparatus and methods - Part 1-5: Radio disturbance and immunity measuring apparatus - Antenna calibration sites and reference test sites for 5 MHz to 18 GHz	EN 55016-1-5	2015
	+ A1	2016		+ A1	2017
	CISPR 16-1-6	2014	Specification for radio disturbance and immunity measuring apparatus and methods - Part 1-6: Radio disturbance and immunity measuring apparatus - EMC antenna calibration	EN 55016-1-6	2015
	+ MD1	2017		+ 1	2017
	+ MD2	2022		+ 2	2022
	CISPR 16-2-3	2016	Specification for radio disturbance and immunity measuring apparatus and methods - Part 2-3: Methods of measurement of disturbances and immunity - Radiated disturbance measurements	EN 55016-2-3	2017
	+ MD1	2019		+ 1	2019
	+ MD2	20— ¹		+ 2	20— ¹ 

¹ Under preparation.

EN IEC 55016-1-4:2019+A2:2023 (E)

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
CISPR TR 16-3	-	Specification for radio disturbance and immunity measuring apparatus and methods -- Part 3: CISPR technical reports		-
CISPR 16-4-2	-	Specification for radio disturbance and immunity measuring apparatus and methods - Part 4-2: Uncertainties, statistics and limit modelling - Measurement instrumentation uncertainty	EN 55016-4-2	-
IEC 60050-191	-	International Electrotechnical Vocabulary - Chapter 191: Dependability and quality of service		-
 IEC 61000-4-20	-	Electromagnetic compatibility (EMC) - Part 4-20: Testing and measurement techniques - Emission and immunity testing in transverse electromagnetic (TEM) waveguides	EN IEC 61000-4-20	-
IEC 61000-4-21	-	Electromagnetic compatibility (EMC) - Part 4-21: Testing and measurement techniques - Reverberation chamber test methods	EN 61000-4-21	- 

CONTENTS

FOREWORD	8
1 Scope	10
2 Normative references	10
3 Terms, definitions and abbreviated terms	11
3.1 Terms and definitions.....	11
3.2 Abbreviated terms.....	15
4 Antennas for measurement of radiated radio disturbance	16
4.1 General.....	16
4.2 Physical parameter (measurand) for radiated disturbance measurements	16
4.3 Antennas for the frequency range 9 kHz to 150 kHz.....	17
4.3.1 General.....	17
4.3.2 Magnetic field antenna.....	17
4.3.3 Shielding of loop antenna	17
4.4 Antennas for the frequency range 150 kHz to 30 MHz.....	17
4.4.1 Electric field antenna	17
4.4.2 Magnetic field antenna.....	18
4.4.3 Balance and electric field discrimination of antennas	18
4.5 Antennas for the frequency range 30 MHz to 1 000 MHz.....	18
4.5.1 General.....	18
4.5.2 Low-uncertainty antenna for use if there is an alleged non-compliance to the electric disturbance field strength limit	18
4.5.3 Antenna characteristics.....	18
4.5.4 Balance of antenna	20
4.5.5 Cross-polar response of antenna	22
4.6 Antennas for the frequency range 1 GHz to 18 GHz.....	23
4.6.1 General.....	23
4.6.2 Receive antenna	23
4.7 Special antenna arrangements – large-loop antenna system.....	25
5 Test sites for measurement of radio disturbance field strength for the frequency range of 9 kHz to 30 MHz	26
6 Test sites for measurement of radio disturbance field strength for the frequency range of 30 MHz to 1 000 MHz	26
6.1 General.....	26
6.2 OATS.....	26
6.2.1 General.....	26
6.2.2 Weather-protection enclosure	26
6.2.3 Obstruction-free area	26
6.2.4 Radio-frequency ambient environment of a test site	28
6.2.5 Ground plane.....	28
6.3 Suitability of other test sites	28
6.3.1 Other ground-plane test sites.....	28
6.3.2 Test sites without ground plane (FAR)	29
6.4 Test site validations	29
6.4.1 General.....	29
6.4.2 Overview of test site validations.....	30
6.5 Basic parameters of the NSA method for OATS and SAC.....	30

6.5.1	General equation and table of theoretical NSA values.....	30
6.5.2	Antenna calibration	34
6.6	Reference site method for OATS and SAC	34
6.6.1	General.....	34
6.6.2	Antennas not permitted for RSM measurements	35
6.6.3	Determination of the antenna pair reference site attenuation on a REFTS	35
6.6.4	Determination of the antenna pair reference site attenuation using an averaging technique on a large OATS.....	36
6.7	Validation of an OATS by the NSA method	39
6.7.1	Discrete frequency method	39
6.7.2	Swept frequency method.....	40
6.8	Validation of a weather-protection-enclosed OATS or a SAC	41
6.9	Possible causes for exceeding site acceptability limits	43
6.10	Site validation for FARs	44
6.10.1	General.....	44
6.10.2	RSM for FAR sites	48
6.10.3	NSA method for FAR sites	50
6.10.4	Site validation criteria for FAR sites	52
6.11	Evaluation of set-up table and antenna tower.....	52
6.11.1	General.....	52
6.11.2	Evaluation procedure for set-up table influences	53
7	Test sites for measurement of radio disturbance field strength for the frequency range 1 GHz to 18 GHz	54
7.1	General.....	54
7.2	Reference test site.....	55
7.3	Test site validation.....	55
7.3.1	General.....	55
7.3.2	Acceptance criterion for site validation.....	56
7.4	Antenna requirements for S_{VSWR} standard test procedure	56
7.4.1	General.....	56
7.4.2	Transmit antenna	57
7.4.3	Antennas and test equipment for the S_{VSWR} reciprocal test procedure	59
7.5	Required positions for site validation testing	60
7.5.1	General.....	60
7.5.2	Descriptions of S_{VSWR} measurement positions in a horizontal plane (Figure 23).....	60
7.5.3	Descriptions of S_{VSWR} additional measurement positions (Figure 24).....	61
7.5.4	Summary of S_{VSWR} measurement positions.....	62
7.6	S_{VSWR} site validation – standard test procedure	65
7.7	S_{VSWR} site validation – reciprocal test procedure using an isotropic field probe	66
7.8	S_{VSWR} conditional measurement position requirements	67
7.9	S_{VSWR} site validation test report.....	68
7.10	Limitations of the S_{VSWR} site validation method.....	68
7.11	Alternative test sites	69
8	Common mode absorption devices	69
8.1	General.....	69
8.2	CMAD S -parameter measurements	69
8.3	CMAD test jig.....	69

8.4	Measurement method using the TRL calibration.....	70
8.5	Specification of ferrite clamp-type CMAD	72
8.6	CMAD performance (degradation) check using spectrum analyzer and tracking generator.....	73
9	Reverberating chamber for total radiated power measurement	75
9.1	General.....	75
9.2	Chamber.....	75
9.2.1	Chamber size and shape	75
9.2.2	Door, openings in walls, and mounting brackets.....	75
9.2.3	Stirrers	76
9.2.4	Test for the efficiency of the stirrers.....	76
9.2.5	Coupling attenuation.....	77
10	TEM cells for immunity to radiated disturbance measurement.....	78
Annex A (normative) Parameters of antennas		79
A.1	General.....	79
A.2	Preferred antennas	79
A.2.1	General.....	79
A.2.2	Calculable antenna	79
A.2.3	Low-uncertainty antennas	79
A.3	Simple dipole antennas.....	80
A.3.1	General.....	80
A.3.2	Tuned dipole.....	81
A.3.3	Shortened dipole.....	81
A.4	Broadband antenna parameters	82
A.4.1	General.....	82
A.4.2	Antenna type	83
A.4.3	Specification of the antenna.....	83
A.4.4	Antenna calibration	84
A.4.5	Antenna user information	84
Annex B (XXX) (Void)		85
Annex C (normative) Large-loop antenna system for magnetic field induced-current measurements in the frequency range of 9 kHz to 30 MHz.....		86
C.1	General.....	86
C.2	Construction of an LLAS	86
C.3	Construction of a large-loop antenna (LLA).....	86
C.4	Validation of an LLAS	91
C.5	Construction of the LLAS verification dipole antenna.....	93
C.6	Conversion factors	94
C.6.1	General.....	94
C.6.2	Current conversion factors for an LLAS with non-standard diameter	95
C.6.3	Conversion of LLAS measured current to magnetic field strength.....	96
C.7	Examples	98
Annex D (normative) Construction details for open area test sites in the frequency range of 30 MHz to 1 000 MHz (see Clause 6).....		100
D.1	General.....	100
D.2	Ground plane construction	100
D.2.1	Material	100
D.2.2	Roughness	100
D.3	Services to EUT.....	101

D.4	Weather-protection enclosure construction	101
D.4.1	Materials and fasteners.....	101
D.4.2	Internal arrangements.....	102
D.4.3	Size	102
D.4.4	Uniformity with time and weather	102
D.5	Turntable and set-up table	102
D.6	Receive antenna mast installation.....	103
Annex E (xxx)	(Void).....	104
Annex F (informative)	Basis for ± 4 dB site acceptability criterion (see Clause 6).....	105
F.1	General.....	105
F.2	Error analysis.....	105
Annex G (informative)	Examples of uncertainty budgets for site validation of a COMTS using RSM with a calibrated antenna pair (see 6.6)	107
G.1	Quantities to be considered for antenna pair reference site attenuation calibration using the averaging technique	107
G.2	Quantities to be considered for antenna pair reference site attenuation calibration using a REFTS	108
G.3	Quantities to be considered for COMTS validation using an antenna pair reference site attenuation	109
Annex H (informative)	Definition of uncertainty in cross-polar response measurement	110
H.1	General.....	110
H.2	Example uncertainty estimate	112
H.3	Rationale for the estimates of input quantities in Table H.1 and Table H.3	113
H.4	Measurement of XPR below 100 MHz at an OATS	114
Bibliography	116
Figure 1	– Schematic of radiation from EUT reaching an LPDA antenna directly and via ground reflection at a 3 m site, showing the beamwidth half-angle, φ , at the reflected ray.....	19
Figure 2	– RX antenna E-plane radiation pattern example, with limit area shaded for 3 m distance and 2 m EUT width.....	24
Figure 3	– Determination of maximum useable EUT width using half-power beamwidth	24
Figure 4	– Determination of maximum useable EUT height using half-power beamwidth	25
Figure 5	– Obstruction-free area of a test site with a turntable	27
Figure 6	– Obstruction-free area with stationary EUT	27
Figure 7	– Test point locations for 3 m and 10 m test distances	36
Figure 8	– Paired test point locations for all test distances.....	38
Figure 9	– Example of paired test point selection for a test distance of 10 m.....	38
Figure 10	– Illustration of an investigation of influence of antenna mast on A_{APR}	39
Figure 11	– Typical antenna positions for a weather-protected OATS or a SAC – vertical polarization validation measurements	42
Figure 12	– Typical antenna positions for a weather-protected OATS or a SAC – horizontal polarization validation measurements	42
Figure 13	– Typical antenna positions for a weather-protected OATS or a SAC – vertical polarization validation measurements for a smaller EUT	43
Figure 14	– Typical antenna positions for a weather-protected OATS or a SAC – horizontal polarization validation measurements for a smaller EUT	43
Figure 15	– Measurement positions for FAR site validation	46

Figure 16 – Example of one measurement position and antenna tilt for FAR site validation	48
Figure 17 – Typical quasi free-space test site reference SA measurement set-up	50
Figure 18 – Theoretical free-space NSA as a function of frequency for different measurement distances [see Equation (16)].....	52
Figure 19 – Position of the antenna relative to the edge above a rectangle set-up table (top view).....	54
Figure 20 – Antenna position above the set-up table (side view).....	54
Figure 21 – Transmit antenna E-plane radiation pattern example (this example is for informative purposes only).....	58
Figure 22 – Transmit antenna H-plane radiation pattern (this example is for informative purposes only).....	59
Figure 23 – S_{VSWR} measurement positions in a horizontal plane (see 7.5.2 for description).....	60
Figure 24 – S_{VSWR} positions (height requirements)	62
Figure 25 – S_{VSWR} conditional measurement position requirements	68
Figure 26 – Definition of the reference planes inside the test jig	70
Figure 27 – The four configurations for the TRL calibration	72
Figure 28 – Limits for the magnitude of S_{11} , measured according to the provisions of 8.1 to 8.3	73
Figure 29 – Example of a 50 Ω adaptor construction in the vertical flange of the jig	74
Figure 30 – Example of a matching adaptor with balun or transformer	74
Figure 31 – Example of a matching adaptor with resistive matching network.....	75
Figure 32 – Example of a typical paddle stirrer	76
Figure 33 – Range of coupling attenuation as a function of frequency for a chamber using the stirrer shown in Figure 32	77
Figure A.1 – Short dipole antenna factors for $R_L = 50 \Omega$	82
Figure C.1 – The LLAS, consisting of three mutually perpendicular large-loop antennas	88
Figure C.2 – An LLA containing two opposite slits, positioned symmetrically with respect to the current probe C	89
Figure C.3 – Construction of an LLA slit.....	89
Figure C.4 – Example of an LLA slit construction using a strap of printed circuit board to obtain a rigid construction	90
Figure C.5 – Construction of the metal box containing the current probe.....	90
Figure C.6 – Example showing the routing of several cables from an EUT to minimize capacitive coupling from the leads to the LLAS	91
Figure C.7 – The eight positions of the LLAS verification dipole during validation of an LLA	92
Figure C.8 – Reference validation factors for loops of 2 m, 3 m, and 4 m diameters.....	92
Figure C.9 – Construction of the LLAS verification dipole antenna	94
Figure C.10 – Sensitivity S_D of an LLA with diameter D relative to an LLA with 2 m diameter	95
Figure C.11 – Conversion factor C_{dA} [for conversion into dB(μ A/m)] for three standard measurement distances d	97
Figure D.1 – The Rayleigh criterion for roughness in the ground plane	101
Table 1 – Site validation methods applicable for OATS, OATS-based, SAC, and FAR site types	30

Table 2 – Theoretical normalized site attenuation, A_N – recommended geometries for broadband antennas ^a (1 of 2)	32
Table 3 – Example template for A_{APR} data sets	35
Table 4 – RSM frequency steps	35
Table 5 – Maximum dimensions of test volume versus test distance	44
Table 6 – Frequency ranges and step sizes for FAR site validation	48
Table 7 – S_{VSWR} measurement position designations (1 of 3)	63
Table 8 – S_{VSWR} reporting requirements	68
Table C.1 – Reference validation factors of Figure C.8 for loops of 2 m, 3 m, and 4 m diameters	93
Table C.2 – Sensitivity S_D of an LLA with diameter D relative to an LLA with 2 m diameter (Figure C.10)	96
Table C.3 – Magnetic field strength conversion factor C_{dA} for three measurement distances (Figure C.11)	98
Table D.1 – Maximum roughness for 3 m, 10 m and 30 m measurement distances	101
Table F.1 – Error budget	105
Table G.1 – Antenna pair reference site attenuation calibration using the large-OATS averaging technique	107
Table G.2 – Antenna pair reference site attenuation calibration using REFTS	108
Table G.3 – COMTS validation using an antenna pair reference site attenuation	109
Table H.1 – Example uncertainty estimate for XPR measurement in a FAR and assumed $a_{xpT} = 22$ dB, $a_{xpR} = 34$ dB	113
Table H.2 – Uncertainties depending on other values of A_{xpT} (other assumptions as in Table H.1)	114
Table H.3 – Example uncertainty estimate for XPR measurement at an OATS and assumed $a_{xpT} = 22$ dB, $a_{xpR} = 34$ dB	115

INTERNATIONAL ELECTROTECHNICAL COMMISSION
INTERNATIONAL SPECIAL COMMITTEE ON RADIO INTERFERENCE

**SPECIFICATION FOR RADIO DISTURBANCE AND
IMMUNITY MEASURING APPARATUS AND METHODS –**

**Part 1-4: Radio disturbance and immunity measuring apparatus –
Antennas and test sites for radiated disturbance measurements**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights

This fourth edition cancels and replaces the third edition published in 2010, Amendment 1:2012 and Amendment 2:2017. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- provisions are added to address test site validation in the frequency range from 30 MHz to 1000 MHz using the reference site method, to take into account the receive antenna radiation pattern in the frequency range from 1 GHz to 18 GHz, and further details on test site validation using the NSA method with broadband antennas in the frequency range from 30 MHz to 1 000 MHz.

CISPR 16-1-4:2019+A1:2020
© IEC 2020

International Standard CISPR 16-1-4 has been prepared by CISPR subcommittee A: Radio-interference measurements and statistical methods.

It has the status of a basic EMC publication in accordance with IEC Guide 107, *Electromagnetic compatibility – Guide to the drafting of electromagnetic compatibility publications*.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
CIS/A/1262/FDIS	CIS/A/1275/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of CISPR 16 series, under the general title *Specification for radio disturbance and immunity measuring apparatus and methods*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

SPECIFICATION FOR RADIO DISTURBANCE AND IMMUNITY MEASURING APPARATUS AND METHODS –

Part 1-4: Radio disturbance and immunity measuring apparatus – Antennas and test sites for radiated disturbance measurements

1 Scope

This part of CISPR 16 specifies the characteristics and performance of equipment for the measurement of radiated disturbances in the frequency range 9 kHz to 18 GHz. Specifications for antennas and test sites are included.

NOTE In accordance with IEC Guide 107, CISPR 16-1-4 is a basic EMC publication for use by product committees of the IEC. As stated in Guide 107, product committees are responsible for determining the applicability of the EMC standard. CISPR and its sub-committees are prepared to cooperate with product committees in the evaluation of the value of particular EMC tests for specific products.

The requirements of this publication apply at all frequencies and for all levels of radiated disturbances within the CISPR indicating range of the measuring equipment.

Methods of measurement are covered in Part 2-3, further information on radio disturbance is given in Part 3, and uncertainties, statistics and limit modelling are covered in Part 4 of CISPR 16.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

CISPR 16-1-1, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-1: Radio disturbance and immunity measuring apparatus – Measuring apparatus*

CISPR 16-1-5:2014, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-5: Radio disturbance and immunity measuring apparatus – Antenna calibration sites and reference test sites for 5 MHz to 18 GHz*
CISPR 16-1-5:2014/AMD1:2016

CISPR 16-1-6:2014, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-6: Radio disturbance and immunity measuring apparatus – EMC antenna calibration*
CISPR 16-1-6:2014/AMD1:2017

CISPR 16-2-3:2016, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 2-3: Methods of measurement of disturbances and immunity – Radiated disturbance measurements*

CISPR TR 16-3, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 3: CISPR technical reports*