



BSI Standards Publication

Geometrical product specifications (GPS) – Acceptance and reverification tests for coordinate measuring systems (CMS)

Part 5: Coordinate measuring machines (CMMs) using single and multiple stylus contacting probing systems using discrete point and/or scanning measuring mode

National foreword

This British Standard is the UK implementation of EN ISO 10360-5:2020. It is identical to ISO 10360-5:2020. It supersedes BS EN ISO 10360-4:2001 and BS EN ISO 10360-5:2010, which are withdrawn.

The UK participation in its preparation was entrusted to Technical Committee TPR/1, Technical Product Realization.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2020
Published by BSI Standards Limited 2020

ISBN 978 0 580 98595 9

ICS 17.040.30; 17.040.40

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 April 2020.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

EUROPEAN STANDARD

EN ISO 10360-5

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2020

ICS 17.040.30

Supersedes EN ISO 10360-4:2000,

English Version

**Geometrical product specifications (GPS) - Acceptance
and reverification tests for coordinate measuring
systems (CMS) - Part 5: Coordinate measuring machines
(CMMs) using single and multiple stylus contacting
probing systems using discrete point and/or scanning
measuring mode (ISO 10360-5:2020)**

Spécification géométrique des produits (GPS) -
Essais de réception et de vérification périodique
des systèmes à mesurer tridimensionnels
(SMT) - Partie 5: MMT utilisant des systèmes
de palpage à stylet simple ou à stylets multiples
utilisant un mode de mesurage par point
discret et/ou par scan (ISO 10360-5:2020)

Geometrische Produktspezifikation (GPS) -
Annahmepfung und Bestätigungsprüfung
für Koordinatenmesssysteme (KMS) - Teil
5: Prüfung der Antastabweichungen von
Koordinatenmessgeräten (KMG) mit berührendem
Messkopfsystem im Einzelpunkt- und/oder
Scanningbetrieb (ISO 10360-5:2020)

This European Standard was approved by CEN on 22 March 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 10360-5:2020) has been prepared by Technical Committee ISO/TC 213 "Dimensional and geometrical product specifications and verification" in collaboration with Technical Committee CEN/TC 290 "Dimensional and geometrical product specification and verification" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2020, and conflicting national standards shall be withdrawn at the latest by October 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10360-4:2000 and EN ISO 10360-5:2010.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 10360-5:2020 has been approved by CEN as EN ISO 10360-5:2020 without any modification.

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Symbols	8
5 Rated operating conditions	10
5.1 Environmental conditions.....	10
5.2 Operating conditions.....	11
6 Acceptance tests and reverification tests	11
6.1 General.....	11
6.2 Measuring equipment.....	12
6.2.1 Test sphere.....	12
6.2.2 Styli specification and information.....	13
6.3 Single-stylus probing test.....	14
6.3.1 Application.....	14
6.3.2 Principle.....	14
6.3.3 Procedure.....	14
6.3.4 Data analysis.....	16
6.4 Scanning mode test.....	16
6.4.1 Principle.....	16
6.4.2 Procedure.....	16
6.4.3 Data analysis.....	18
6.5 Multi-stylus test: Fixed multi-probe and multi-stylus probing systems.....	19
6.5.1 Principle.....	19
6.5.2 Procedure.....	19
6.5.3 Data analysis.....	22
6.6 Multi-stylus test: Articulating probing systems.....	22
6.6.1 Principle.....	22
6.6.2 Procedure.....	23
6.6.3 Data analysis.....	25
6.7 Data analysis for multi-stylus tests.....	25
6.7.1 Location error.....	25
6.7.2 Opposing-styli projected location error.....	25
6.7.3 Multi-stylus size and form error.....	25
7 Conformance with specification: Acceptance and reverification tests	26
8 Applications	27
8.1 Acceptance tests.....	27
8.2 Reverification tests.....	27
8.3 Interim checks.....	27
Annex A (informative) Ring gauge tests	28
Annex B (informative) Checking the probing system prior to the ISO 10360-2 test	32
Annex C (informative) Interpretation of multi-stylus test results	33
Annex D (normative) Maximum permissible error/limit specification methods	35
Annex E (informative) Workpiece related influences	37
Annex F (normative) Acceptance tests and reverification tests using small sphere test equipment	39
Annex G (informative) Relation to the GPS matrix model	40

Bibliography **42**

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 213, *Dimensional and geometrical product specifications and verification* in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 290, *Dimensional and geometrical product specification and verification*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 10360-5:2010), which has been technically revised.

It also incorporates with a technical revision the tests contained within ISO 10360-4:2000 and, as such, it cancels and replaces ISO 10360-4:2000.

The main changes to the previous edition are as follows:

- the adoption of new symbology;
- the addition of an optional ring gauge test;
- changes to acceptable test parameters e.g. test sphere diameter;
- changes to Location results evaluation including an “opposing styli” evaluation.

A list of all parts in the ISO 10360 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document is a geometrical product specification (GPS) standard and is to be regarded as a general GPS standard (see ISO 14638). It influences chain link F of the chains of standards on size, distance, form, orientation, location and run-out.

The ISO GPS matrix model given in ISO 14638 gives an overview of the ISO GPS system of which this document is a part. The fundamental rules of ISO GPS given in ISO 8015 apply to this document and the default decision rules given in ISO 14253-1 apply to specifications made in accordance with this document, unless otherwise indicated.

For more detailed information about the relation of this document to other standards and the GPS matrix model see [Annex G](#).

The acceptance and reverification tests described in this document are applicable to coordinate measuring machines (CMMs) that use contacting probes, with or without multiple styli or multiple articulated-probe positions, when measuring using discrete point and/or scanning mode.

Experience has shown that the multi-stylus errors calculated using this document are significant and, at times, represent the dominant errors in the CMM. Owing to the virtually infinite variety of modern CMM probing system configurations, the description of the tests specified by this document provides a testing protocol for specification, but the actual test coverage has been limited to provide a practical subset of tests which are intended to reveal typical errors associated with probing configurations in a limited amount of time. The tests are intended to provide information on the ability of a CMM to measure a feature or features using a contacting probe and, when relevant, using multiple styli, multiple probes or multiple articulated-probe positions.

The situations to which they are applicable include:

- single-stylus probing systems;
- multiple styli connected to the CMM probe (e.g. a star);
- installations using an articulating probing system (motorized or manual) that can be prequalified;
- installations using a repeatable probe-changing system;
- installations using a repeatable stylus-changing system;
- installations including a scanning probe, capable of being used in a scanning mode;
- multi-probe installations.

It is believed that the procedures given in this document will be helpful in identifying CMM system uncertainty components for specific measurement tasks, and that the user will be able to reduce errors by removing contributing elements such as long probe extensions and styli, and then by retesting the new configuration set.

The tests in this document are sensitive to many errors, attributable to both the CMM and the probing system, and are intended to be performed in addition to the length-measuring tests given in ISO 10360-2.

The primary objective is to determine the practical performance of the complete CMM and probing system. Therefore, the tests are designed to reveal measuring errors which are likely to occur when such a combined system is used on real workpieces, for example errors generated by the interaction between large probe-tip-offset lengths and uncorrected CMM rotation errors. The errors found here differ from those found in the EL tests in ISO 10360-2, because with multiple styli the net CMM travel may be very different from the measured length. See [Annex C](#) for more information.

This document complements ISO 10360-7 (CMMs equipped with imaging probing systems), ISO 10360-8 (CMMs with optical distance sensors), ISO 10360-9 (CMMs with multiple probing systems) and ISO 10360-2 (CMMs used for measuring linear dimensions).

Geometrical product specifications (GPS) – Acceptance and reverification tests for coordinate measuring systems (CMS) —

Part 5:

Coordinate measuring machines (CMMs) using single and multiple stylus contacting probing systems using discrete point and/or scanning measuring mode

1 Scope

This document specifies acceptance and periodic reverification tests of CMM performance with contacting probing systems and is only applicable to CMMs using:

- any type of contacting probing system; and
- spherical or hemispherical stylus tip(s).

NOTE CMM probing performance tests are specified by the maximum permissible errors (MPEs), due to the impracticality of isolating the performance of the probing system from that of the CMM, even on a small artefact such as a test sphere.

This document applies to CMMs supplied with any of the following:

- a) single-stylus probing systems;
- b) multi-stylus probing systems with fixed multiple styli attached to a single probe (e.g. “star” stylus);
- c) multiple probing systems such as those with a stylus for each of their probes;
- d) systems with articulating probing systems;
- e) stylus and probe changing systems;
- f) manual (non-driven) and automated CMMs;
- g) installations including a scanning probe, capable of being used in a scanning mode.

This document is not applicable to non-contacting probing systems, which require different testing procedures.

The term ‘combined CMM and multi-stylus probing system size error’ has been shortened to ‘multi-stylus size error’ for convenience. This applies in similar cases.

If it is desirable to isolate the probing system performance as far as is practical, the influence of the CMM can be minimized but not eliminated. See [Annex C](#) for more information.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.