Guidelines for the
Evaluation of Uncertainty
of Test Values Associated
With the Verification of
Dimensional Measuring
Instruments to Their
Performance Specifications

AN AMERICAN NATIONAL STANDARD



Guidelines for the
Evaluation of Uncertainty
of Test Values Associated
With the Verification of
Dimensional Measuring
Instruments to Their
Performance Specifications

AN AMERICAN NATIONAL STANDARD



Date of Issuance: June 12, 2020

This Standard will be revised when the Society approves the issuance of a new edition.

ASME issues written replies to inquiries concerning interpretations of technical aspects of this Standard. Interpretations are published on the Committee web page and under http://go.asme.org/InterpsDatabase. Periodically certain actions of the ASME B89 Committee may be published as Cases. Cases are published on the ASME website under the B89 Committee Page at http://go.asme.org/B89committee as they are issued.

Errata to codes and standards may be posted on the ASME website under the Committee Pages to provide corrections to incorrectly published items, or to correct typographical or grammatical errors in codes and standards. Such errata shall be used on the date posted.

The B89 Committee Page can be found at http://go.asme.org/B89committee. There is an option available to automatically receive an e-mail notification when errata are posted to a particular code or standard. This option can be found on the appropriate Committee Page after selecting "Errata" in the "Publication Information" section.

ASME is the registered trademark of The American Society of Mechanical Engineers.

This code or standard was developed under procedures accredited as meeting the criteria for American National Standards. The Standards Committee that approved the code or standard was balanced to assure that individuals from competent and concerned interests have had an opportunity to participate. The proposed code or standard was made available for public review and comment that provides an opportunity for additional public input from industry, academia, regulatory agencies, and the public-at-large.

ASME does not "approve," "rate," or "endorse" any item, construction, proprietary device, or activity.

ASME does not take any position with respect to the validity of any patent rights asserted in connection with any items mentioned in this document, and does not undertake to insure anyone utilizing a standard against liability for infringement of any applicable letters patent, nor assume any such liability. Users of a code or standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, is entirely their own responsibility.

Participation by federal agency representative(s) or person(s) affiliated with industry is not to be interpreted as government or industry endorsement of this code or standard.

ASME accepts responsibility for only those interpretations of this document issued in accordance with the established ASME procedures and policies, which precludes the issuance of interpretations by individuals.

No part of this document may be reproduced in any form, in an electronic retrieval system or otherwise, without the prior written permission of the publisher.

The American Society of Mechanical Engineers Two Park Avenue, New York, NY 10016-5990

Copyright © 2020 by
THE AMERICAN SOCIETY OF MECHANICAL ENGINEERS
All rights reserved
Printed in U.S.A.

CONTENTS

Foreword	d	iv
Committe	ee Roster	V
Correspo	er	
Preface .		viii
1	Scope	1
2	Definitions	1
3	References	1
4	Basic Concepts of Test Values and Test Value Uncertainty	2
5	Evaluation of Test Value Uncertainty	5
Nonman	datory Appendix	
Α	Example: Verification of a CMM to Its ASME B89.4.10360.2 Performance Specifications	9
Figures		
4.1-1	A Caliper Under Verification to Its MPE Specification by a Verification System Composed of Calibrated Gauge Blocks Using a Simple Acceptance Decision Rule	3
5.1-1	An Example of Five Test Values, All From the Same Reference Length	5
Table		
A-1-1	Test Value Uncertainty for Test Values Obtained Using a 500-mm Calibrated Gauge Block and a Test	10