



**BSI Standards Publication**

## **Dentistry — Corrosion test methods for dental amalgam**

---

## National foreword

This Published Document is the UK implementation of ISO/TS 17988:2020. It supersedes PD ISO/TS 17988:2014, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/106/1, Dental restorative and orthodontic materials.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2020  
Published by BSI Standards Limited 2020

ISBN 978 0 539 00673 5

ICS 11.060.10

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 March 2020.

### Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

---

# TECHNICAL SPECIFICATION

**ISO/TS  
17988**

Second edition  
2020-03

---

---

## Dentistry — Corrosion test methods for dental amalgam

*Médecine bucco-dentaire — Essais de corrosion des amalgames  
dentaires*



Reference number  
ISO/TS 17988:2020(E)

© ISO 2020



## **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b>	<b>v</b>
<b>Introduction</b>	<b>vii</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>1</b>
<b>4 Sampling</b>	<b>3</b>
<b>5 Preparation of dental amalgam test-pieces</b>	<b>4</b>
5.1 General	4
5.1.1 Temperature	4
5.1.2 Mixing	4
5.2 Cylindrical test-pieces for use in the immersion and potentiostatic corrosion test procedures	4
5.2.1 Mass of dental amalgam to be mixed	4
5.2.2 Apparatus for the preparation of dental amalgam cylindrical test-pieces	4
5.2.3 Packing	8
5.3 Disc-shaped test-pieces for use in the Hertzian-loading strength-reduction test	9
5.3.1 Apparatus for the preparation of dental amalgam disc-shaped test-pieces	9
5.3.2 Materials and tolerances for construction of the mould	9
5.3.3 Packing the mould, removal of test-piece and inspection for surface defects	10
<b>6 Determination of the resistance to corrosion by the immersion procedure</b>	<b>10</b>
6.1 Principle	10
6.2 Reagents for the test solution and cleaning the apparatus	10
6.3 Apparatus	13
6.4 Mercury vapour analyser requirements	14
6.5 Cleaning the glassware	15
6.6 Assembly of the immersion corrosion test apparatus	15
6.7 Test-piece production	15
6.8 Preparation of the 0,1 mol/l lactic acid solution	16
6.9 Finishing the dental amalgam test-piece	16
6.10 Test procedure	16
6.10.1 First determination	16
6.10.2 Second determination	17
6.11 Analysis to determine the metal ion and mercury vapour release	18
6.12 Test report	18
<b>7 Determination of the corrosion by the potentiostatic procedure</b>	<b>19</b>
7.1 Principle	19
7.2 Test-piece preparation	19
7.3 Corrosion test cell requirements	20
7.3.1 Corrosion cell	20
7.3.2 Temperature control	20
7.3.3 Volume of the electrolyte	20
7.4 Reference electrode probe requirements	20
7.4.1 Reference electrode and its control	20
7.4.2 Temperature of the reference electrode	20
7.4.3 Positioning of the reference electrode	20
7.5 Potentiostat requirements	21
7.6 Reagents	21
7.7 Preparation of the electrolyte	21
7.8 Test procedure	21
7.9 Data acquisition and processing	21
7.9.1 General	21
7.9.2 Computer-controlled potentiostat	22

7.9.3	Coulometer .....	22
7.9.4	Data-logging and integration .....	22
7.10	Calculation of the total charge transported .....	22
7.10.1	Test-pieces embedded by casting without masking .....	22
7.10.2	Test-pieces embedded by casting with masking .....	22
7.11	Test report .....	22
<b>8</b>	<b>Determination of the resistance to corrosion by the Hertzian-loading strength-reduction test .....</b>	<b>23</b>
8.1	Principle .....	23
8.2	Test solution (artificial saliva) .....	23
8.2.1	Reagents .....	23
8.2.2	Stock solutions .....	24
8.2.3	Test solution (artificial saliva) .....	24
8.3	Test-piece production and procedure for test-piece conditioning .....	24
8.3.1	Apparatus .....	24
8.3.2	Control test-pieces .....	25
8.3.3	Corrosion test-pieces .....	25
8.3.4	Replacement test-pieces .....	26
8.4	Mechanical testing .....	26
8.4.1	Apparatus for mechanical testing .....	26
8.4.2	Procedure .....	26
8.5	Treatment of data .....	28
8.6	Test report .....	28
	<b>Bibliography .....</b>	<b>30</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 1, *Filling and restorative materials*.

This second edition cancels and replaces the first edition (ISO/TS 17988:2014), which has been technically revised. The main changes to the previous edition are as follows:

- The scope has been extended to include products that are within the scope of ISO 20749.
- [Clause 3](#) includes additional terms and definitions.
- [Clause 4](#): quantities required for the production of test-pieces for each of the three test procedures are given now as the mass of dental amalgam alloy per test-piece, in place of the total mass of dental amalgam alloy for the complete test (i.e. the estimated quantity for all test-pieces including permitted replacements).
- [5.2.2.2](#) and [5.3.2](#): the parameter  $R_a$  has replaced  $R_k$  to specify surface roughness on steel moulds.
- [5.3.2](#): the surface roughness of the tapered hole in the Hertzian-indentation strength-reduction test-piece mould has been revised.
- [8.3.1.1](#) and [8.3.1.2](#): two additional items have been added to the list of apparatus.
- [8.3.1.4](#): blood dilution vials without protuberances on the interior base surface might not be available. A means by which the required flat surface can be created has been added.
- [8.3.4](#) and [8.4.2.2](#): a technical addition has been made to the procedure. Instructions are given for replacing test-pieces from which invalid results had been produced. Also, advice is given to make the maximum number of permitted replacements at the time that the actual test-pieces are made (to avoid a possible 31-day delay should a result be invalid and a replacement test-piece be required).
- [8.4.2.2](#): instructions are given to inspect the substrate disc and to replace it if damage is observed.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).



## Introduction

This document gives the practical details of three test methods for the measurement of the resistance to corrosion of dental amalgam. These corrosion test methods are laboratory procedures for evaluating the relative performances of dental amalgam alloy products. They are designed to produce a measurable effect (and differences between products) within a relatively short time period, a time period appropriate for a comparative laboratory evaluation.

The results of these tests should not be used for any biocompatibility claims, for which their use is inappropriate.

Should other corrosion test procedures emerge in time as suitable for application in comparative evaluations of dental amalgam products, they will be included in future editions of this document.

# Dentistry — Corrosion test methods for dental amalgam

## 1 Scope

This document gives the details of test procedures for evaluating the corrosion resistance of dental amalgam formed from products that are within the scopes of ISO 24234 and ISO 20749.

This document is not applicable to other metallic materials in which an alloy powder reacts with a liquid alloy to produce a solid metallic material intended for dental restoration.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 286-2, *Geometrical product specifications (GPS) — ISO code system for tolerances on linear sizes — Part 2: Tables of standard tolerance classes and limit deviations for holes and shafts*

ISO 1942, *Dentistry — Vocabulary*

ISO 3585, *Borosilicate glass 3.3 — Properties*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 4287, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Terms, definitions and surface texture parameters*

ISO 6344-1, *Coated abrasives — Grain size analysis — Part 1: Grain size distribution test*

ISO 7488, *Dentistry — Mixing machines for dental amalgam*

ISO 13897, *Dentistry — Dental amalgam reusable mixing-capsules*

ISO 24234, *Dentistry — Dental amalgam*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

### 3.1

#### **dental amalgam alloy**

alloy in fine particles, composed mainly of silver, tin and copper, which when mixed with dental mercury produces a dental amalgam for dental restoration

[SOURCE: ISO 20749:2017, 3.1]