



BSI Standards Publication

**Microbeam analysis — Electron probe
microanalysis — Guidelines for the specification
of certified reference materials (CRMs)**

National foreword

This British Standard is the UK implementation of ISO 14595:2023. It supersedes BS ISO 14595:2014, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CII/9, Microbeam analysis.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 16320 9

ICS 71.040.30; 71.040.99

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2023.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

**Microbeam analysis — Electron probe
microanalysis — Guidelines for the
specification of certified reference
materials (CRMs)**

*Analyse par microfaisceaux — Microanalyse par sonde à électrons —
Lignes directrices pour les spécifications des matériaux de référence
certifiés (CRM)*





COPYRIGHT PROTECTED DOCUMENT

© ISO 2023, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Preparation of the research material	2
4.1 Selection of material.....	2
4.2 Preliminary inspection of the material.....	2
5 Heterogeneity of material	2
5.1 Sample preparation.....	2
5.2 Sample size.....	3
5.3 Test conditions.....	3
5.4 Test procedure.....	4
5.5 Statistical evaluation of data.....	4
5.6 Criteria for certification.....	8
6 Stability of the research material	8
7 Determination of the chemical composition of CRMs	9
7.1 Classification of CRMs.....	9
7.2 Determination of classification of CRMs.....	9
7.3 Selection of analytical method.....	9
7.4 CRM material tested by EPMA only.....	9
8 CRM specimen preparation, packaging, transportation, and storage	9
8.1 Preparation of CRM specimen.....	9
8.2 Packaging.....	10
8.3 Storage.....	10
8.4 Repolishing and recoating of CRMs.....	10
9 CRM certificate	10
9.1 Classification of CRM.....	10
9.2 Contents of the certificate.....	10
Annex A (informative) Spreadsheet instructions for the statistical evaluation of heterogeneity data	11
Annex B (normative) Suggested classification of CRMs for EPMA	14
Annex C (informative) Example of a certificate for EPMA CRMs	15
Bibliography	16

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 202, *Microbeam analysis*, Subcommittee SC 2, *Electron probe microanalysis*.

This third edition cancels and replaces the second edition (ISO 14595:2014), which has been technically revised.

The main changes are as follows:

- formulae to calculate the uncertainty of the mean mass fraction in 5.6 has been modified by adding a power of 1/2;
- the requirement of the dead time for the energy dispersive spectrometer test has been removed;
- editorial changes have been made to improve the consistency of terms and definitions used throughout the document.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

For electron probe microanalysis (EPMA), which is a comparative quantitative analytical method used globally, certified reference materials (CRMs) play a crucial role in the analytical accuracy.

This document has been developed to facilitate international exchange and compatibility of analysis data in EPMA.

It aims to give guidance on evaluating and selecting reference materials (RMs), on evaluating the extent of heterogeneity and stability of RMs. It gives recommendations for the determination of the chemical composition of RMs for production as EPMA-certified reference materials.

Microbeam analysis — Electron probe microanalysis — Guidelines for the specification of certified reference materials (CRMs)

1 Scope

This document specifies recommendations for single-phase certified reference materials (CRMs) used in electron probe microanalysis (EPMA). It also provides guidance on the use of CRMs for the microanalysis of flat, polished specimens. It does not cover organic or biological materials.

This document supplements ISO 17034. A producer of CRM must also comply with ISO 17034. In case of conflict, ISO 17034 takes precedence.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

heterogeneity

measured variation in compositions of elements measured from a group of specimens

Note 1 to entry: The contributions to heterogeneity include the uncertainties in the measurements from specimen to specimen, from micrometre to micrometre within each specimen, and from the test procedure itself.

3.2

research material

material that appears to have the physical and chemical characteristics required of a CRM, but which is to be examined in detail, including the determination of chemical composition, stability, and micro-heterogeneity and macro-heterogeneity, before certification as a CRM

3.3

stability

<general>resistance of a specimen to chemical and physical change during long-term storage at normal temperature and pressure

3.4

stability

<EPMA>resistance of the material to changes in chemical composition during electron bombardment, i.e. the resistance to change of the intensity of the relevant characteristic X-rays observed during the time the specimen is exposed to the electron beam