

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Medical electrical equipment –
Part 2-47: Particular requirements for the basic safety and essential performance
of ambulatory electrocardiographic systems**

**Appareils électromédicaux –
Partie 2-47: Exigences particulières pour la sécurité de base et les performances
essentielles des systèmes d'électrocardiographie ambulatoires**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2012 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

Useful links:

IEC publications search - www.iec.ch/searchpub

The advanced search enables you to find IEC publications by a variety of criteria (reference number, text, technical committee,...).

It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available on-line and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 30 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary (IEV) on-line.

Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Liens utiles:

Recherche de publications CEI - www.iec.ch/searchpub

La recherche avancée vous permet de trouver des publications CEI en utilisant différents critères (numéro de référence, texte, comité d'études,...).

Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

Just Published CEI - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications de la CEI. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 30 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (VEI) en ligne.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.



IEC 60601-2-47

Edition 2.0 2012-02

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –

Part 2-47: Particular requirements for the basic safety and essential performance of ambulatory electrocardiographic systems

Appareils électromédicaux –

Partie 2-47: Exigences particulières pour la sécurité de base et les performances essentielles des systèmes d'électrocardiographie ambulatoires

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE **XB**
CODE PRIX

ICS 11.040.55

ISBN 978-2-88912-911-9

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
201.1 Scope, object and related standards.....	7
201.2 Normative references	9
201.3 Terms and definitions	9
201.4 General requirements.....	11
201.5 General requirements for testing of ME EQUIPMENT.....	12
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	12
201.7 ME EQUIPMENT identification, marking and documents.....	12
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	13
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	14
201.10 Protection against unwanted and excessive radiation HAZARDS.....	14
201.11 Protection against excessive temperatures and other HAZARDS.....	14
201.12 Accuracy of controls and instruments and protection against hazardous outputs	14
201.13 HAZARDOUS SITUATIONS and fault conditions	38
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	38
201.15 Construction of ME EQUIPMENT	38
201.16 ME SYSTEMS.....	39
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	39
202 Electromagnetic compatibility – Requirements and tests	40
Annexes	42
Annex AA (informative) Particular guidance and rationale	43
Bibliography.....	64
Index of defined terms used in this particular standard.....	65
Figure 201.101 – General test circuit for 201.12.4.4.....	29
Figure 201.102 – Test signal for input dynamic range test according to 201.12.4.4.101	30
Figure 201.103 – Test circuit for common mode rejection according to 201.12.4.4.103	33
Figure 201.104 – Test circuit for pacemaker pulse tolerance according to 201.12.4.4.109.....	37
Figure 202.101 – Test set-up for conductive emission test according to 202.6.1.1.2 and radiated emission and radiated immunity test according to 202.6.1.1.2 and 202.6.2.3.2.....	41
Table 201.101 – Distributed additional ESSENTIAL PERFORMANCE requirements	11
Table 201.102 – LEAD WIRE colour codes	13
Table 201.103 – Requirements for all arrhythmia algorithms	17
Table 201.104 – Requirements for algorithms with optional capabilities	18
Table 201.105 – Beat label classifications	22
Table 201.106 – Example of noise floor calculation results	24
Table 201.107 – Example of HRV test results	25
Table 201.108 – Run sensitivity summary matrix	25
Table 201.109 – Run positive predictivity summary matrix	26

Table AA.1 – Records to be included in a complete test..... 44

Table AA.2 – Example of a line-format, beat-by-beat performance report..... 48

Table AA.3 – Condensed beat-by-beat summary matrix containing 11 elements 49

Table AA.4 – Summary table (matrix format) of beat-by-beat comparison 49

Table AA.5 – Example of a line-format SHUTDOWN report..... 50

Table AA.6 – Example of a line-format report..... 51

Table AA.7 – Example of VF performance report 51

Table AA.8 – Example of false VF performance report..... 51

Table AA.9 – Example of a line-format couplet and run performance report 52

Table AA.10 – Example of device measurements of synthetic test patterns..... 53

Table AA.11 – Example of predicted ideal values for synthetic test patterns 54

Table AA.12 – Example of choice of test patterns 54

Table AA.13 – Example of RMS interval differences 57

Table AA.14 – Example of summary of frequency components 58

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-47: Particular requirements for the basic safety and essential performance of ambulatory electrocardiographic systems

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-47 has been prepared by subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 2001. It constitutes a technical revision. This edition was revised to align structurally with the 2005 edition of IEC 60601-1.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/963/FDIS	62D/980/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This particular standard concerns the basic safety and essential performance of AMBULATORY ELECTROCARDIOGRAPHIC SYSTEMS. It amends and supplements IEC 60601-1 (third edition 2005): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard. The requirements of this particular standard take priority over those of the general standard.

A “General guidance and rationale” for the requirements of this particular standard is included in Annex AA.

It is considered that a knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this standard.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-47: Particular requirements for the basic safety and essential performance of ambulatory electrocardiographic systems

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of AMBULATORY ELECTROCARDIOGRAPHIC SYSTEMS, hereafter referred to as ME SYSTEMS

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

Within the scope of this standard are systems of the following types:

- a) systems that provide continuous recording and continuous analysis of the ECG allowing full re-analysis giving essentially similar results. The systems may first record and store the ECG and analyse it later on a separate unit, or record and analyse the ECG simultaneously. The type of storage media used is irrelevant with regard to this standard;
- b) systems that provide continuous analysis and only partial or limited recording not allowing a full re-analysis of the ECG.

The safety aspects of this standard apply to all types of systems falling in one of the above-mentioned categories.

If the AMBULATORY ELECTROCARDIOGRAPHIC SYSTEM offers automatic ECG analysis, minimal performance requirements for measurement and analysis functions apply. MEDICAL ELECTRICAL EQUIPMENT covered by IEC 60601-2-25 and IEC 60601-2-27 are excluded from the scope of this standard.

This standard does not apply to systems that do not continuously record and analyse the ECG (for example, 'intermittent event recorders').

201.1.2 Object

Replacement:

¹ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.