
Washer-disinfectors —

Part 7:

**Requirements and tests for washer-
disinfectors employing chemical
disinfection for non-invasive, non-
critical thermolabile medical devices
and healthcare equipment**

Laveurs désinfecteurs —

*Partie 7: Exigences et essais pour les laveurs désinfecteurs utilisant la
désinfection chimique pour les dispositifs médicaux et les équipements
de soins thermosensibles non invasifs et non critiques*



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Performance requirements	2
4.1 General.....	2
4.2 Cleaning.....	3
4.3 Disinfection.....	3
4.4 Final rinsing.....	5
4.5 Self-disinfection.....	5
4.6 Drying.....	6
4.7 Water treatment equipment.....	6
4.7.1 General.....	6
4.7.2 Disinfection of water treatment equipment.....	6
4.7.3 Maintenance of piping.....	7
5 Mechanical requirements	7
5.1 Materials — Design, manufacture, and assembly.....	7
5.2 Process verification.....	7
6 Testing for conformity	7
6.1 General.....	7
6.2 Test load.....	8
6.2.1 Loading with standard goods.....	8
6.2.2 Loading with special goods.....	8
6.3 Water used for rinsing following disinfection.....	8
6.4 Load dryness.....	8
6.4.1 General.....	8
6.4.2 Procedure.....	8
6.4.3 Results.....	8
6.5 Thermometric tests.....	8
6.5.1 General.....	8
6.5.2 Load temperature test.....	8
6.6 Chemical dosing tests.....	9
6.6.1 General.....	9
6.6.2 Reused process chemicals.....	9
6.7 Tests of cleaning efficacy.....	9
6.7.1 General.....	9
6.7.2 Materials.....	9
6.7.3 Procedure.....	10
6.7.4 Results.....	10
6.8 Test of disinfection efficacy.....	10
6.8.1 General.....	10
6.8.2 Preliminary tests on chemical disinfectants.....	10
6.8.3 Self-disinfection tests.....	11
6.8.4 Chemical disinfection of the load.....	12
7 Documentation	12
8 Information to be provided by the manufacturer	12
9 Marking, labelling, and packaging	13
10 Information to be requested from the purchaser by the manufacturer	13
Annex A (normative) Summary of test programmes	14

Annex B (normative) Methods for microbiological evaluation of disinfection of liquid transport system	16
Annex C (normative) Tests for microbiological contamination of post-disinfection rinse water	21
Annex D (normative) Preparation and evaluation of indicators for microbiological testing of the efficacy of chemical disinfection of the load	22
Annex E (informative) Examples of test locations for the tests with biological indicators	26
Bibliography	30

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

ISO 15883-7 was prepared by the European Committee for Standardization (CEN) Technical Committee CEN/TC 102, *Sterilizers for medical purposes*, in collaboration with Technical Committee ISO/TC 198, *Sterilization of health care products*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

ISO 15883 consists of the following parts, under the general title *Washer-disinfectors*:

- *Part 1: General requirements, terms and definitions and tests*
- *Part 2: Requirements and tests for washer-disinfectors employing thermal disinfection for surgical instruments, anaesthetic equipment, bowls, dishes, receivers, utensils, glassware, etc.*
- *Part 3: Requirements and tests for washer-disinfectors employing thermal disinfection for human waste containers*
- *Part 4: Requirements and tests for washer-disinfectors employing chemical disinfection for thermolabile endoscopes*
- *Part 5: Test soils and methods for demonstrating cleaning efficacy* [Technical Specification]
- *Part 6: Requirements and tests for washer-disinfectors employing thermal disinfection for non-invasive, non-critical medical devices and healthcare equipment*
- *Part 7: Requirements and tests for washer-disinfectors employing chemical disinfection for non-invasive, non-critical thermolabile medical devices and healthcare equipment*

Introduction

It is intended that this introduction is to be read in conjunction with the introduction to ISO 15883-1.

This part of ISO 15883 is the seventh of a series specifying the performance of washer-disinfectors. It specifies the particular requirements for performance applicable to washer-disinfectors employing chemical disinfection for non-invasive, non-critical thermolabile medical devices, and healthcare equipment. Its requirements apply to washer-disinfectors used for cleaning and disinfection of thermolabile equipment for use without further treatment in healthcare settings. Such reusable equipment needs to be cleaned and disinfected, but processing in a washer-disinfector for surgical instruments (see ISO 15883-2), for human waste containers (see ISO 15883-3), for endoscopes (see ISO 15883-4), or for non-invasive, non-critical medical devices, and healthcare equipment employing thermal disinfection (see ISO 15883-6) is inappropriate and/or impractical. Examples of such equipment are bedsteads and bedside furniture, trolleys and transport carts, operating tables, footwear, wheelchairs, or aids for the disabled.

Requirements for washer-disinfectors for other applications are specified in other parts of ISO 15883.

In respect to any potential adverse effects on the quality of water intended for human consumption caused by use of the washer-disinfector, it is noteworthy that

- a) until verifiable international criteria are adopted, the existing national regulations concerning the use and/or characteristics of the washer-disinfector remain in force (e.g. the requirement of backflow prevention), and
- b) the ISO 15883 series of standards provides no information as to whether the washer-disinfector may be used without restriction in any of the ISO member states.

Washer-disinfectors —

Part 7:

Requirements and tests for washer-disinfectors employing chemical disinfection for non-invasive, non-critical thermolabile medical devices and healthcare equipment

1 Scope

This part of ISO 15883 specifies the particular requirements for washer-disinfectors (WD) intended to be used for the cleaning and chemical disinfection, in a single operating cycle, of reusable items such as the following:

- a) bedframes;
- b) bedside tables;
- c) transport carts;
- d) containers;
- e) surgical tables;
- f) sterilization containers;
- g) surgical clogs;
- h) wheelchairs, aids for the disabled.

This part of ISO 15883 also specifies the performance requirements for the cleaning and disinfection of the washer-disinfector and its components and accessories which may be necessary in order to achieve the required performance.

Devices identified within the scopes of ISO 15883-2, ISO 15883-3, ISO 15883-4, and ISO 15883-6 do not fall within the scope of this part of ISO 15883.

In addition, the methods are specified, as well as instrumentation and instructions required for type testing, works testing, validation (installation, operation, and performance qualification on first installation), routine control, and monitoring, as well as requalifications required to be carried out periodically and after essential repairs.

NOTE WDs corresponding to this part of ISO 15883 can also be used for cleaning and chemical disinfection of other thermolabile and reusable devices as recommended by the device manufacturer.

The performance requirements specified in this part of ISO 15883 may not ensure the inactivation or removal of the causative agent(s) (prion proteins) of Transmissible Spongiform Encephalopathies.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.