



CSA C22.2 No. 61010-2-091:19

Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 2-091: Particular requirements for cabinet X-ray systems
(IEC 61010-2-091:2019, MOD)

CSA C22.2 n° 61010-2-091:19

Règles de sécurité pour appareils électriques de mesure, de régulation et de laboratoire — Partie 2-091 : Exigences particulières pour les équipements à rayons X montés en armoire
(IEC 61010-2-091:2019, MOD)



Standards Update Service

CSA C22.2 No. 61010-2-091:19 November 2019

Title: *Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 2-091: Particular requirements for cabinet X-ray systems*

To register for e-mail notification about any updates to this publication

- go to store.csagroup.org
- click on **Product Updates**

The **List ID** that you will need to register for updates to this publication is **2427716**.

If you require assistance, please e-mail techsupport@csagroup.org or call 416-747-2233.

Visit CSA Group's policy on privacy at www.csagroup.org/legal to find out how we protect your personal information.

CONTENTS

FOREWORD	3
INTRODUCTION	6
1 Scope and object	7
2 Normative references	8
3 Terms and definitions	8
4 Tests	9
5 Marking and documentation	10
6 Protection against electric shock	13
7 Protection against mechanical HAZARDS	15
8 Resistance to mechanical stresses	16
9 Protection against the spread of fire	16
10 Equipment temperature limits and resistance to heat	16
11 Protection against HAZARDS from fluids and solid foreign objects	16
12 Protection against radiation, including laser sources, and against sonic and ultrasonic pressure	16
13 Protection against liberated gases and substances, explosion and implosion	19
14 Components and subassemblies	19
15 Protection by INTERLOCKS	20
16 HAZARDS resulting from application	22
17 RISK assessment	22
Annexes	23
Annex L (informative) Index of defined terms	24
Annex AA (normative) Standard phantom for PARTIALLY PROTECTED EQUIPMENT	25
Annex BB (informative) Dose limit recommendation for occupational radiation	26
Bibliography	27
Figure AA.1 – Phantom for measurement of PARTIALLY PROTECTED EQUIPMENT	25
Table 101 – Durations of high voltage strength test	14
Table BB.1 – Dose limits	26

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 2-091: Particular requirements for cabinet X-ray systems

1 Scope and object

This clause of Part 1 is applicable, except as follows:

1.1 Scope

1.1.1 Equipment included in scope

Deletion:

Delete the first paragraph.

Replacement:

Replace the second paragraph (above items a) to c)) with the following new text:

This part of IEC 61010 specifies particular safety requirements for cabinet X-ray systems, which fall under any of categories a), b) or c) below.

Addition:

Add the two following new paragraphs at the end of the subclause:

Equipment covered by this document can be both PROTECTED EQUIPMENT or PARTIALLY PROTECTED EQUIPMENT, with X-ray generator voltage up to 500 kV.

A cabinet X-ray system is a system that contains an X-ray tube installed in a cabinet, which, independently of existing architectural structures except the floor on which it may be placed, is intended to contain at least that portion of a material being irradiated, provide radiation attenuation and prevent operator access to the radiation beam, during generation of X-radiation.

These cabinet X-ray systems are used in industrial, commercial, and public environments, for example, to inspect materials, to analyse materials, and to screen baggage.

1.1.2 Equipment excluded from scope

Addition:

Add the following new items to the list:

- aa) Equipment intended to apply X-radiation to humans or animals;
- bb) Equipment incorporating an X-ray tube but not incorporating complete shielding against X-radiation HAZARDS, such as:
 - equipment intended to be used within a shielded room which excludes personnel during operation;
 - equipment intended to be used with separate portable or temporary shielding;
 - equipment intended to produce an emerging beam of X-radiation.